

Temperature Monitoring & Data Logger CDC/VFC Compliance Table



CDC Requirements (for VFC vaccines or vaccines purchased with public funds)	Basic	Loggers	Loggers	Logger/Wi-Fi
Traceable® certificate supplied (traceable to NIST)	✓	✓	✓	✓
Calibration report included	✓	✓	✓	✓
Annual recalibration by ILAC/MRA accredited laboratory available	✓	✓	✓	✓
CDC Recommendations*				
Detachable probe in a thermal buffered material	✓	✓	✓	✓
Calibration certificate includes methods, procedures, uncertainties & results	✓	✓	✓	✓
Can display current temperature and minimum/maximum memory	✓	✓	✓	✓
Minimum/maximum memory reset	✓	✓	✓	✓
Alarm for out-of-range temperatures	✓	✓	✓	✓
Digital display easily readable from outside the storage unit	✓	✓	✓	✓
Low-battery indicator	✓	✓	✓	✓
Minimum uncertainty of ±0.5°C (±1°F)		✓	✓	✓
Continuous monitoring and active display		✓	✓	✓
User programmable logging intervals		✓	✓	✓
Data can be downloaded using software or retrieved from a website		✓	✓	✓
Data will not overwrite when full			✓	✓
Capable of storing 4,000 readings		✓	✓	

*Each individual state/local health department immunization program may have specific requirements. Consult your immunization program for more information. Recommendations and requirements are subject to change without prior notice.

[Go to Web Page >>](#)



info@VeriCorMed.com
(608) 526-6901

For more information & pricing: www.VeriCorMed.com

